

INSPECTION REPORT

Item	Portable Combustible Gas Detector		Model	BX616	
Quantity	1 pc		Date	11-01-2018	
Batch number	C104029001				
Apperance	Clean, No mechanical defacement				
Main Function	LCD display Audible, visible and vibration alarming Real-time display of time and remaining voltage Settable H-alarm and L-alarm levels				
Detecting Gas	H ₂ S	CO	CH ₄		O ₂
Detecting Range	0~100ppm	0~1000ppm	0~100%LEL		0~30%VOL
Requested display	50ppm ±5ppm	500ppm ±50ppm	20%LEL ±5%LEL	50%LEL ±5%LEL	Within 3min, the figure changes less than 0.50%vol
Testing result	Qualification				
Response Time	<30s	<30s	<30s		<30s
Testing condition	Temperature: 20°C Humidity: 40%RH				
Testing Result	Qualification				

Inspector: Zhenzhen Wang 

QC Supervisor: Linfeng Gao

Contact :

EASTERN ENERGY CO.,LTD. (HEAD OFFICE)

40/4 Vitoondumri Rd., Banbueng, Banbueng, Chonburi 20170

Tel: 0-3844-6117, Fax: 0-3844-6200

Email: info@ete.co.th / sale@ete.co.th

Web: www.eastern-energy.com



**EASTERN
ENERGY**