

WBGT Meter with Color-Coded LCD



Takes the guesswork out of heat stress calculations.

This unique heat stress meter features bright multi-color LCD backlights corresponding to OSHA's heat stress alert levels. This visual alert system can be seen at a distance allowing management and workers to be instantly aware when the heat index enters potentially dangerous zones. These features, combined with the optional audible alarm help to mitigate the risk of heat-related injuries in the workplace. Calculates heat stress values for direct sunlight and shade and displays humidity, maximum and minimum ambient (TA) and black globe (TG) temperatures. Also indicates low battery and a hold function. Comes complete with clear instructions, integrated tripod screw and 4 AAA batteries. N.I.S.T. Traceable Certificate of Compliance available (for RH & Air Temperature only).

DIM: $12'' \times 3'' \times 1\frac{1}{2}''$ (302 × 75 × 37 mm). WEIGHT: 10.6 oz (300 g)

Screen

Gray

235339

Green

248688

Orange

30.

388|788

Red

Red Flashing

Workforce Precautions

Basic heat safety training

Implement precautions

Additional precautions

are necessary to

protect workers.

Analysis of activity

protect workers.

Triggers even

more aggressive

protective measures.

level is necessary to

and heighten awareness.

and planning.

WBGT

< 80.0°F

< 26.6°C

80.1 to 84.9°F

26.7 to 29.3°C

85.0 to 87.9°F

29.4 to 31.0°C

88.0 to 89.9°F

31.1 to 32.1°C

> 90.0°F

> 32.2°C

No.	Description
801038	WBGT Meter with Color-Coded LCD
801038C	WBGT Meter with Color-Coded
	LCD Certified

	Range	Resolution	Accuracy
Ambient Air Temperature	0 to 50°C 32 to 122°F	0.1	±0.6°C (1.5°F)
Globe Temperature	0 to 80°C 32 to 176°F	0.1	±1°C @15 to 40°C ±2°C outside this range
RH	5 to 90%	0.1	±4%
WBGT	0 to 50°C 32 to 122°F	0.1	±1°C @15 to 40°C ±2°C outside this range





Contact:

EASTERN ENERGY CO.,LTD. (HEAD OFFICE)

40/4 Vitoondumri Rd., Banbueng, Banbueng, Chonburi 20170 Tel: 0-3844-6117 Fax: 0-3844-6200 Email: info@ete.co.th

Source: Manual of Naval Preventive Medicine (NAVMED P-5010) and OSHA Section 5(a)(1)