



TEST REPORT

1 of 1

MODEL: DT-9850M Particle Counter

Serial Number:180702204

Report Number:201821206

Referance Instrument: FLUKE985

Instrument Series Number:1210993188

Environment Temperature:23±3℃

Humidity: 50±20%RH

Calibration Date:2018-7-9

Issue Date:2018-7-9

Calibrated By: *Li Menglong*

Issue Name: Peng Xingen

Calibrate Type	Display Value	Test Value	Result
Repeatability Testing	≤10%FS	<10%FS	Pass
Distribute Deviation of Particle	±30%	<30%	Pass
Particle Density Value Deviation	±30%FS	<30%FS	Pass
HCHO Measure	±5%F.S	±5%F.S	Pass
Air Flow	2.83L±0.5L	2.90L	Pass

Conclusion: Pass at Calibration Item.

Contact :

EASTERN ENERGY CO.,LTD. (HEAD OFFICE)

40/4 Vitoondumri Rd., Banbueng, Banbueng, Chonburi 20170

Tel: 0-3844-6117 Fax: 0-3844-6200

Email: info@ete.co.th



www.ete.co.th