

TEST REPORT

1 of 1

MODEL:DT-9883M Particle Counter

Serial Number:190202533

Report Number:20196229

Referance Instrument: FLUKE985

Instrument Series Number:1210993188

Environment Temperature:23±3℃

Humidity: 50±20%RH

Calibration Date:2019-3-4

Issue Date:2019-3-4

Calibrated By *Li Menglong*

Issue Name: Peng Xingen

Calibrate Type	Display Value	Test Value	Result
Repeatability Testing	≤10%FS	<10%FS	Passed
Distribute Deviation of Particle	±30%	<30%	Passed
Particle Density Value Deviation	±30%FS	<30%FS	Passed
Air Flow	2.83L±5%	2.80L	Passed
HCHO Measure	±5%F.S	±5%F.S	Passed
TVOC Measure	±5%F.S	±5%F.S	Passed
CO ₂ Measure	±5%±75ppm	±5%±75ppm	Passed

Conclusion: Pass at Calibration Item.

Contact :**EASTERN ENERGY CO.,LTD. (HEAD OFFICE)**40/4 Vitoondumri Rd., Banbueng, Banbueng,
Chonburi 20170

Tel: 0-3844-6117 Fax: 0-3844-6200

Email: info@ete.co.th

**EASTERN
ENERGY**www.ete.co.th