

Vibration Meter



Model: VM-6360 (Enhanced Type)
 VM-6320 (Functional Type)
 VM-6310 (Basic Type)

Applications

Used for measuring periodic motion, to check the imbalance and deflecting of moving machinery. Specifically designed for present measuring various mechanical vibration. So as to provide the data for the quality control, run time and equipment upkeep.

- * VM-6360 has wide frequency range (10Hz~10kHz) in acceleration mode
- * VM-6320 / VM-6310 are simple and easy to use.



Features

- * In accordance with ISO 2954, used for periodic measurements, to detect out-of-balance, misalignment and other mechanical faults in rotating machines.
- * Specially designed for easy on site vibration measurement of all rotating machinery for quality control, commissioning, and predictive maintenance purposes.
- * Individual high quality accelerometer for accurate and repeatable measurements.
- * Optional headphones for use as electronic stethoscope.
- * Use RS-232 data output to connect with PC.
- * Provide Bluetooth data output choice.

Vibration Standard

| ISO/IS2373 Motor Quality Standard According As Vibration Velocity | | | | |
|---|-----------|---------------------------------------|-----------|-----------|
| Quality Rank | Rev (rpm) | H: high of shaft (mm) | | |
| | | Maximum vibration velocity rms (mm/s) | | |
| | | 80<H<132 | 132<H<225 | 225<H<400 |
| Normal | 600~3600 | 1.8 | 2.8 | 4.5 |
| Good (R) | 600~1800 | 0.71 | 1.12 | 1.8 |
| | 1800~3600 | 1.12 | 1.8 | 2.8 |
| Excellent (S) | 600~1800 | 0.45 | 0.71 | 1.12 |
| | 1800~3600 | 0.71 | 1.12 | 1.8 |

Contact :

EASTERN ENERGY CO.,LTD. (HEAD OFFICE)

40/4 Vitoondumri Rd., Banbueng, Banbueng,

Chonburi 20170

Tel: 0-3844-6117 Fax: 0-3844-6200

Email: info@ete.co.th






www.ete.co.th



Specifications

| Model | | VM-6360 | VM-6320 | VM-6310 |
|----------------------|--------------|--|--|--------------------------|
| Sensor | | Piezoelectric Transducer | | |
| Measuring Range | Acceleration | 0.1~400 m/s ² Equivalent Peak | 0.1~199.9 m/s ² Equivalent Peak | — |
| | Velocity | 0.01~400 mm/s True RMS | 0.01~199.9 mm/s True RMS | 0.01~199.9 mm/s True RMS |
| | Displacement | 0.001~4.0 mm Equivalent Peak-peak | 0.001~1.999 mm Equivalent Peak-peak | — |
| Frequency Range | Acceleration | 10Hz~10kHz | | — |
| | Velocity | 10Hz~1kHz | | 10Hz~1kHz |
| | Displacement | 10Hz~1kHz | | — |
| Accuracy | | 5% of Reading + 2 digits | | |
| Operating Conditions | Temperature | 0~50 °C | | |
| | Humidity | <90 %RH | | |
| Power Supply | | 4x1.5V AAA (UM-4) Battery | 4x1.5V AA (UM-3) Battery | |
| Dimensions | | 124x62x30mm | 160x68x32mm | |
| Weight | | 120 g (Not Including Batteries) | 181 g (Not Including Batteries) | |
| Standard Accessories | | Main Unit | | |
| | | Piezoelectric Transducer | | |
| | | Powerful Rare Earth Magnet | | |
| | | Probe (Cone) & Probe (Spherical) | | |
| | | Carrying Case (B04) | | |
| | | Manual Book | | |
| | | Headset | | |
| Optional Accessories | | RS-232C Data Cable with Software | | |
| | | Bluetooth Data Adapter with Software | | |

Accessories

| Accessories | Diagram | Using Situations | Using Method |
|-----------------------------|---|---|---|
| Piezoelectric Transducer |  | General vibration parameters measurement of objects. | Be used with Powerful Rare Earth Magnet & Stinger Probe. |
| Rare Earth Magnet |  | Magnetic objects with flat surface, roughness of less than Ra1.6, acceleration ≤ 20m/s. | connect the vibration sensor with Rare Earth Magnet with the M5 bolt included. And then place the Rare Earth Magnet to the object to be tested. |
| Stinger Probe (Ball / Cone) |  | Frequency is less than 1KHz and vibration energy is not small. | Connect the needle to the sensor directly by using probe groupware. |

Contact :

EASTERN ENERGY CO.,LTD. (HEAD OFFICE)

40/4 Vitoondumri Rd., Banbueng, Banbueng,

Chonburi 20170

Tel: 0-3844-6117 Fax: 0-3844-6200

Email: info@ete.co.th

