



SHENZHEN EVERBEST MACHINERY INDUSTRY CO.,LTD

19TH BUILDING,5TH REGION,BAIWANGXIN INDUSTRIAL PARK.SONGBAI RD,BAIMANG,XILI,NANSHAN SHENZHEN CHINA 518108
TEL:86.755.27353188 FAX:86.755.27652253

TEST REPORT

1 of 1

MODEL: DT-9880 Particle Counter User Manual

Serial Number:161208722

Report Number:201631895

Reference Instrument: FLUKE985

Instrument Series Number:1210993188

Environment Temperature:23±3°C

Humidity: 50±20%RH

Calibration Date:2016-12-29

Issue Date:2016-12-29

Calibrated By: *Li Menglong*

Issue Name: Peng Xingen

Calibrate Type	Display Value	Test Value	Result
Repeatability Testing	≤10%FS	<10%FS	Pass
Distribute Deviation of Particle	±30%	<30%	Pass
Particle Density Value Deviation	±30%FS	<30%FS	Pass
Air Flow	2.83L±5%	2.70L	Pass

Conclusion: Pass at Calibration Item.

Eastern Energy Co., Ltd.
40/4 Vitoondumri Rd., Banbueng
Banbueng, Chonburi, Thailand



EASTERN ENERGY

Tel: 66-3844-6117
sale@ete.co.th
www.eastern-energy.com

深圳華盛昌機械實業有限公司

深圳市南山區西麗白芒鬆白路
百旺信工業區5區19棟 518108

電話: 86.755.27353188
傳真: 86.755.27652253