SMART SENSOR

CERTIFICATE OF CALIBRATION

Instruments details	
Model number:	AS8901
Description:	Gas Monitor
Serial number:	3422477
Date of manufacture:	2018-10-18

Reference documents for the calibration: JJG 915-2008, JJG 365-2008
JJG 695-2003, JJG 693-2011

Place and environmental conditions of the calibration:

Place	Temperature	RH
SMART SENSOR FQC	23℃～27℃	50%RH～75%RH

Function Range Mode	Normalized Value	Actual Value	Permissible Error
Measure value	(O_2)	(O_2)	(O_2)
氧气 (O_2)	20.9	20.8	-0.1

Contact :
EASTERN ENERGY CO.,LTD. (HEAD OFFICE)
40/4 Vitoondumri Rd., Banbueng, Banbueng, Chonburi 20170
Tel: 0-3844-6117 Fax: 0-3844-6200
Email: info@ete.co.th
www.ete.co.th

www.eastern-energy.com